



WWW.SECURITYSYSTEMS.NO

MARCA BLANCA BLACKBODY-TH



Blackbody - Calibration device for thermographic cameras - Infrared emission of 35°C ~ 60°C - Stability $\pm 0.1\text{--}0.2^\circ\text{C/h}$ - Emissivity 0.97 ± 0.02 - Guarantees a measurement accuracy of $\pm 0.3^\circ\text{C}$

- Blackbody
- Guarantees a measurement accuracy of $\pm 0.3^\circ\text{C}$
- Temperature range: 35°C ~ 60°C
- Accuracy $\pm 0.2^\circ\text{C}$
- Radiation area 78x78 mm
- Stability $\pm 0.1\text{--}0.2^\circ\text{C/h}$
- Emissivity 0.97 ± 0.02